



校准证书

CALIBRATION CERTIFICATE

证书编号

Certificate No.

DAFTJAA058950003



第 1 页, 共 3 页

Page 1 of 3 Pages

委托方

Client

首都医科大学宣武医院(首都医科大学宣武医院GCP药房)

联络信息

Information

北京市西城区长椿街45号

仪器名称

Description

温控探头

型号规格

Model/Type

T41P

制造厂商

Manufacturer

智筑行方

出厂编号

Serial No.

442522000415

管理号

Asset No.

样品接收日期

Date of Receipt

2026-06-18

校准日期

Date of Calibration

2026-06-18

发布日期

Issued Date

2026-06-18

批准人:

Approved Signatory

周庆南

审核:

Inspected by

廖志勇

校准:

Calibrated by

王希飞



校准专用章
校准专用章

(stamp)



扫一扫验真伪

本实验室地址: 广东省深圳市南山区桃源街道珠光社区珠光路珠光创新科技园 3 栋三楼、五楼, 4 栋 103

The laboratory address: 3&5/F., Building 3, and Room 103, Building 4, Zhuguang Chuangxin Science & Technology Park, Zhuguang Road, Taoyuan Street, Nanshan District, Shenzhen, Guangdong, China

电话 (Tel): 0755-86070926

邮政编码 (Post Code): 518055

传真 (Fax): 0755-26645455

电子邮件 (E-mail): daftest@szdaf.com

网址 (Web): <https://www.szdaf.com/>



校准说明

DIRECTIONS OF CALIBRATION

证书编号：DAFTJAA058950003

Certificate No.

第 2 页，共 3 页

Page 2 of 3 pages

- 1.本结果仅对当次被测样品有效，如有疑问请在15个工作日反馈。
(The results is ONLY valid for the tested sample, please feedback to us within 15 working days if you have any question.)
- 2.本证书校准结果均可溯源至国际单位制（SI）单位和社会公用计量标准。
(The calibration results of this certificate can be traced back to the International System of Units (SI) units the social public measurement standards.)
- 3.本证书编号具有唯一性，若替换证书，自发出后原证书即可作废。
(Each certificate has a unique number. If replaced the certificate, the original certificate will be invalid once the new certificate number is issued.)
- 4.证书中如有最大允许误差、判定结果，仅供参考，其中“P”代表“Pass”，“F”代表“Fail”。
(In the datasheet, MPE & determination is only for reference, "P" represents "Pass" and "F" represents "Fail".)
- 5.本证书/报告提供的结果仅对本次被检的计量器具负责。
(The results presented in this document apply only to the verified instrument.)
- 6.本证书未经本机构书面授权，不得部分复制。
(The Certificate shall not be partly reproduced without written approval of the laboratory.)
- 7.本次校准的技术依据：
(Reference documents for Calibration:)
JJF(苏)95-2024《数字温度计校准规范》

8.本次校准使用的主要测量标准 (Main Standards of Measurement Used in the Calibration):

名称 Description	编号 Serial No.	溯源机构名称 Traceability Institution	证书编号 Certificate No.	有效期 Due Date	技术特征 Technique Character
标准铂电阻温度计	210340	华南国家计量测试中心	RZD202440356	2026-08-15	二等
智能多通道测温仪	DTMC-2301	中国计量科学研究院	DCzd2026-00209	2027-01-15	相对误差 $\leq 1 \times 10^{-5}$

9.校准地点、环境条件 (Place and environmental conditions of the calibration)

地点： 本公司热学室

温度： 21.2 °C

相对湿度： 42 %

Place

Temperature

Relative Humidity

10.建议复校时间间隔： 12 个月，送检单位可根据实际情况自主决定。

Suggested calibration interval is 12 months or it can be altered depending on the actual usage of the user.



校准结果

Result of Calibration

报告编号: DAFTJAA058950003
Report No.

第 3 页, 共 3 页
Page 3 of 3 Pages

1、外观及功能检查: 正常
Appearance and Function Check: Normal

2、校准结果:
Calibration Result:

校准点(°C) Calibration Point	示值误差(°C) Indication Error	允许误差(°C) MPE	结论(Pass/Fail) Conclusion
2	-0.1	± 1.0	P
5	-0.1	± 1.0	P
8	-0.2	± 1.0	P

备注:

Notes:

1、本次测量结果扩展不确定度(Expanded Uncertainty of The Measurement Results):

温度(Temperature): $U = 0.2 \text{ }^{\circ}\text{C}$, $k=2$

2、依据(Reference Document): JJF 1059.1-2012 测量不确定度评定与表示

JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement

(以下空白)

(The Below is Blank)

